

Page 6 of 11

**WADSWORTH CENTER
NEW YORK STATE DEPARTMENT OF HEALTH
ENVIRONMENTAL LABORATORY APPROVAL PROGRAM**

Proficiency Test Report

Lab Id: 11642 USGS NATIONAL WATER QUALITY LABORATORY
 EPA Lab Code: POB 25046-MS 407-BLDG 95
 CO00009 DENVER, CO 80225-0046
 (303) 236-2000
 Director: MR. MERLE SHOCKEY

Shipment Date : 24-Jul-2006
 Closing Date : 31-Aug-2006
 Score Date : 20-Sep-2006

This report may contain data that are not covered by the NVLAP accreditation.

** Indicates NVLAP accredited analyte evaluated using the USEPA's National Standards for Water Proficiency Testing Studies Criteria Document.
 NVLAP Lab Code 200387-0. ELAP is an A2LA accredited Proficiency Testing Provider. Certificate Number 1785.01

Shipment: 295 Non Potable Water Chemistry

Analyte Name	Units	Sample ID	Method	Result	Mean/Target *	Warning Limits	Acceptance Limits	Score
Sample: Non Potable Water Base/Neutrals								
2,4-Dinitrotoluene EPA Code: N/A	ug/L	9536	USGS O-3116-87	< 0.43	< 20.0 *			Satisfactory 64 passed out of 68 reported results.
2,6-Dinitrotoluene EPA Code: N/A	ug/L	9536	USGS O-3116-87	137	140		70.7 - 210	Satisfactory 63 passed out of 67 reported results.
2-Chloronaphthalene EPA Code: N/A	ug/L	9536	USGS O-3116-87	68.9	65.6		26.3 - 105	Satisfactory 67 passed out of 69 reported results.
4-Chlorophenylphenyl ether EPA Code: N/A	ug/L	9536	USGS O-3116-87	< 0.34	< 25.0 *			Satisfactory 68 passed out of 69 reported results.
4-Bromophenylphenyl ether EPA Code: N/A	ug/L	9536	USGS O-3116-87	56.5	61.2		24.3 - 98.2	Satisfactory 66 passed out of 67 reported results.
Acenaphthene EPA Code: N/A	ug/L	9536	USGS O-3116-87	54.8	56.0		28.2 - 83.7	Satisfactory 73 passed out of 76 reported results.
Anthracene EPA Code: N/A	ug/L	9536	USGS O-3116-87	18.0	26.9		13.6 - 40.3	Satisfactory 71 passed out of 76 reported results.
Acenaphthylene EPA Code: N/A	ug/L	9536	USGS O-3116-87	68.8	93.4		46.3 - 140	Satisfactory 70 passed out of 75 reported results.
Bis(2-chloroethyl)ether EPA Code: N/A	ug/L	9536	USGS O-3116-87	67.0	65.4		23.9 - 107	Satisfactory 68 passed out of 69 reported results.
Bis(2-chloroethoxy)methane EPA Code: N/A	ug/L	9536	USGS O-3116-87	136	135		67.8 - 202	Satisfactory 65 passed out of 69 reported results.
Bis(2-chloroisopropyl)ether EPA Code: N/A	ug/L	9536	USGS O-3116-87	< 0.38	< 30.0 *			Satisfactory 68 passed out of 69 reported results.
Bis(2-ethylhexyl)phthalate EPA Code: N/A	ug/L	9536	USGS O-3116-87	85.4	119		42.5 - 196	Satisfactory 72 passed out of 73 reported results.